
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002465
14:22 28/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:10
STERILIZING 4:20
136.3 C 222kPa 4:20
136.2 C 223kPa 4:50
136.1 C 222kPa 5:20
136.1 C 223kPa 5:50
136.3 C 225kPa 6:20
136.4 C 225kPa 6:50
136.3 C 224kPa 7:20
136.4 C 225kPa 7:50
136.4 C 225kPa 7:50
VENTING 7:50
AIR DRYING 8:43
CYCLE COMPLETE 38:43

Digital Signature #
42687A82C057243A
