
Midmark M11 - v1.0.5
Total Cycles: 3300

Name: RITTER2
Sterilizer ID: V2313228
01/15/2024 03:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.1 F 0.0
2:00 108.5 F 0.2
4:00 130.3 F 1.3
6:00 159.7 F 4.2
8:00 181.1 F 7.7
10:00 210.2 F 11.4
12:00 217.3 F 3.9
14:00 226.5 F 7.6
16:00 237.6 F 11.6
18:00 232.2 F 8.0
20:00 244.3 F 13.4
22:00 243.1 F 12.1
24:00 254.8 F 18.1
26:00 265.2 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.8 F 29.5
2:00 272.1 F 28.9
3:00 271.3 F 28.9
4:00 272.7 F 29.5

Min 271.2 F 28.6
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.2 F 0.0
5:00 221.5 F 0.0
10:00 215.4 F 0.0
15:00 186.2 F 0.0
20:00 181.6 F 0.0
25:00 179.5 F 0.0
30:00 178.5 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 27:57
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:06:32
CYCLE COMPLETE

01/15/2024 04:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____