
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001549
12:06 30/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:49
STERILIZING 3:46
136.0 C 220kPa 3:46
136.2 C 224kPa 4:17
136.1 C 222kPa 4:47
136.2 C 223kPa 5:17
136.4 C 224kPa 5:47
136.4 C 224kPa 6:17
136.4 C 225kPa 6:47
136.4 C 225kPa 7:17
136.4 C 225kPa 7:17
VENTING 7:17
AIR DRYING 8:07
CYCLE COMPLETE 8:40

Digital Signature #
05D0D638A2478331
