
Midmark M11 - v1.0.0
Total Cycles: 4409

Name: RITTER 1
Sterilizer ID: V2178198
03/27/2024 09:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.6 F 0.0
2:00 134.9 F 1.3
4:00 147.9 F 3.8
6:00 171.0 F 7.1
8:00 193.3 F 10.8
10:00 215.9 F 4.3
12:00 221.9 F 7.2
14:00 230.5 F 11.2
16:00 229.9 F 6.9
18:00 242.0 F 12.2
20:00 239.0 F 9.9
22:00 250.9 F 15.7
24:00 261.2 F 21.6
26:00 270.9 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.8 F 29.0
2:00 271.8 F 29.0
3:00 272.3 F 28.9
4:00 272.1 F 29.2

Min 271.3 F 28.6
Max 273.1 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.2 F 0.0
5:00 217.9 F 0.0
10:00 207.4 F 0.0
15:00 180.8 F 0.0
20:00 177.7 F 0.0
25:00 175.8 F 0.0
30:00 174.7 F 0.0

Drying Complete

FILLING: 0:57
HEATING: 27:30
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:06:13
CYCLE COMPLETE

03/27/2024 10:21 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____