
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002007
14:40 10/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:05
STERILIZING 4:14
136.5 C 225kPa 4:14
136.4 C 226kPa 4:44
136.2 C 224kPa 5:14
136.4 C 226kPa 5:44
136.2 C 225kPa 6:14
136.4 C 225kPa 6:44
136.4 C 226kPa 7:14
136.4 C 226kPa 7:44
136.4 C 226kPa 7:44
VENTING 7:44
AIR DRYING 8:42
CYCLE COMPLETE 28:38

Digital Signature #
D6C279FB1EF9A213
