
Midmark M11 - v1.0.5
Total Cycles: 3893

Name: RITTER2
Sterilizer ID: V2313228
08/26/2024 03:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 97.1 F 0.0
2:00 97.4 F 0.0
4:00 114.8 F 0.1
6:00 149.6 F 1.8
8:00 168.6 F 4.7
10:00 198.0 F 8.1
12:00 202.5 F 12.0
14:00 218.8 F 4.9
16:00 225.4 F 9.0
18:00 238.4 F 13.4
20:00 237.1 F 10.4
22:00 236.6 F 9.0
24:00 247.8 F 14.5
26:00 258.5 F 20.4
28:00 268.3 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.5 F 29.4
2:00 271.2 F 29.2
3:00 272.4 F 29.4
4:00 271.2 F 29.1

Min 271.1 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.2 F 0.0
5:00 230.8 F 0.0
10:00 183.9 F 0.0
15:00 180.4 F 0.0
20:00 178.1 F 0.0
25:00 176.2 F 0.0
30:00 175.5 F 0.0
Drying Complete

FILLING: 0:41
HEATING: 29:17
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:07:39
CYCLE COMPLETE

08/26/2024 04:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____