
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000489
8:34 21/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:09
PRESSURIZING 4:10
STERILIZING 5:11
136.0 C 220kPa 5:11
136.4 C 224kPa 5:41
136.1 C 221kPa 6:11
136.3 C 224kPa 6:41
136.4 C 224kPa 7:11
136.4 C 220kPa 7:41
136.4 C 221kPa 8:11
136.2 C 219kPa 8:41
136.2 C 219kPa 8:41
VENTING 8:41
AIR DRYING 9:33
CYCLE COMPLETE 9:39

Digital Signature #
9861A9100797D4DF
