
Midmark M11 - v1.0.0
Total Cycles: 3407

Name: RITTER 1
Sterilizer ID: V2178198
06/19/2023 02:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.2 F 0.0
2:00 206.2 F 3.6
4:00 216.8 F 8.4
6:00 235.5 F 10.7
8:00 234.9 F 9.8
10:00 238.9 F 8.5
12:00 245.8 F 13.0
14:00 245.3 F 12.6
16:00 258.0 F 19.5
18:00 269.3 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.2
1:00 272.0 F 28.3
2:00 272.7 F 28.8
3:00 272.0 F 28.6
4:00 272.0 F 28.3

Min 271.4 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.1 F 0.0
5:00 231.2 F 0.0
10:00 225.1 F 0.0
15:00 225.2 F 0.0
20:00 225.4 F 0.0
25:00 224.5 F 0.0
30:00 223.2 F 0.0

Drying Complete

FILLING: 2:19
HEATING: 18:58
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:28
CYCLE COMPLETE

06/19/2023 03:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____