
Midmark M11 - v1.0.0
Total Cycles: 2298

Name: RITTER 1
Sterilizer ID: V2178198
09/07/2022 01:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.8 F 0.0
2:00 153.0 F 0.4
4:00 171.4 F 1.5
6:00 197.7 F 5.5
8:00 214.0 F 10.1
10:00 224.6 F 7.5
12:00 232.9 F 10.0
14:00 235.1 F 8.6
16:00 244.3 F 12.9
18:00 242.4 F 11.4
20:00 255.2 F 18.0
22:00 266.4 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 28.9
2:00 271.6 F 28.2
3:00 272.7 F 29.0
4:00 272.1 F 28.4

Min 271.5 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.0 F 0.0
5:00 216.9 F 0.0
10:00 231.9 F 0.0
15:00 228.4 F 0.0
20:00 223.9 F 0.0
25:00 221.8 F 0.0
30:00 218.4 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 23:38
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:01:53
CYCLE COMPLETE

09/07/2022 02:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____