

-----  
Midmark M11 - v1.0.5  
Total Cycles: 497

Name: RITTER2  
Sterilizer ID: V2313228  
10/28/2021 10:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 172.5 F 0.0  
2:00 175.6 F 3.2  
4:00 187.4 F 8.0  
6:00 207.1 F 9.9  
8:00 225.6 F 8.8  
10:00 234.7 F 10.5  
12:00 242.1 F 12.8  
14:00 242.8 F 12.2  
16:00 256.4 F 19.4  
18:00 268.3 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.8 F 29.9  
2:00 272.2 F 29.2  
3:00 271.6 F 29.2  
4:00 272.7 F 30.0  
  
Min 271.5 F 28.9  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.9 F 0.0  
5:00 228.1 F 0.0  
10:00 236.9 F 0.0  
15:00 225.6 F 0.0  
20:00 224.1 F 0.0  
25:00 216.1 F 0.0  
30:00 209.2 F 0.0  
  
Drying Complete

FILLING: 1:50  
HEATING: 19:14  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:58:10  
CYCLE COMPLETE

10/28/2021 12:10 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02