
Midmark M11 - v1.0.5
Total Cycles: 3199

Name: RITTER2
Sterilizer ID: V2313228
12/08/2023 03:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.1 F 0.0
2:00 110.4 F 0.3
4:00 123.5 F 1.9
6:00 151.9 F 4.9
8:00 175.2 F 8.3
10:00 193.7 F 12.1
12:00 212.8 F 5.4
14:00 224.3 F 9.4
16:00 232.7 F 11.6
18:00 236.5 F 10.6
20:00 235.2 F 8.5
22:00 247.2 F 14.6
24:00 258.4 F 20.8
26:00 268.7 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.6
1:00 272.0 F 29.6
2:00 272.0 F 29.6
3:00 272.2 F 29.5
4:00 272.1 F 29.7

Min 271.8 F 29.5
Max 272.5 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.5 F 0.0
5:00 222.9 F 0.0
10:00 197.3 F 0.0
15:00 193.6 F 0.0
20:00 192.7 F 0.0
25:00 191.9 F 0.0
30:00 190.8 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 27:54
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:06:28
CYCLE COMPLETE

12/08/2023 04:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____