
Midmark M11 - v1.0.5
Total Cycles: 4241

Name: RITTER2
Sterilizer ID: V2313228
01/03/2025 10:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.8 F 0.0
2:00 72.0 F 0.0
4:00 81.3 F 0.0
6:00 113.4 F 0.1
8:00 156.2 F 0.4
10:00 177.0 F 3.4
12:00 191.3 F 6.7
14:00 205.1 F 10.2
16:00 216.4 F 8.5
18:00 224.9 F 7.7
20:00 235.7 F 12.0
22:00 234.7 F 9.0
24:00 240.8 F 10.8
26:00 245.9 F 13.5
28:00 257.4 F 19.7
30:00 267.8 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.7 F 29.3
2:00 271.6 F 28.9
3:00 272.6 F 29.3
4:00 271.6 F 28.8

Min 271.4 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.0 F 0.0
5:00 201.6 F 0.0
10:00 213.1 F 0.0
15:00 216.1 F 0.0
20:00 215.7 F 0.0
25:00 216.8 F 0.0
30:00 215.9 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:21
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:09:38
CYCLE COMPLETE

01/03/2025 11:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____