
Midmark M11 - v1.0.5
Total Cycles: 3928

Name: RITTER2
Sterilizer ID: V2313228
09/11/2024 10:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.0 F 0.0
2:00 152.0 F 0.2
4:00 152.6 F 2.8
6:00 179.3 F 6.5
8:00 196.0 F 10.4
10:00 219.7 F 7.3
12:00 225.3 F 8.9
14:00 234.6 F 13.4
16:00 236.4 F 10.4
18:00 236.0 F 8.8
20:00 247.0 F 14.4
22:00 257.7 F 20.2
24:00 267.5 F 26.4
26:00 273.0 F 30.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.8
1:00 272.9 F 30.2
2:00 271.9 F 30.0
3:00 272.7 F 30.1
4:00 271.9 F 30.0

Min 271.7 F 29.7
Max 273.0 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 217.5 F 0.0
10:00 214.8 F 0.0
15:00 208.7 F 0.0
20:00 216.0 F 0.0
25:00 222.5 F 0.0
30:00 220.6 F 0.0

Drying Complete

FILLING: 0:60
HEATING: 26:24
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:05:04
CYCLE COMPLETE

09/11/2024 11:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____