
Midmark M11 - v1.0.5
Total Cycles: 18

Name: RITTER2
Sterilizer ID: V2313228
04/26/2021 11:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.8 F 0.0
2:00 175.8 F 4.1
4:00 198.0 F 9.5
6:00 221.9 F 6.7
8:00 232.1 F 11.4
10:00 236.4 F 10.0
12:00 237.4 F 9.2
14:00 252.1 F 16.8
16:00 264.2 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.5 F 29.7
2:00 271.6 F 28.8
3:00 272.4 F 29.9
4:00 272.5 F 29.3

Min 271.3 F 28.8
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.0 F 0.0
5:00 204.2 F 0.0
10:00 204.2 F 0.0
15:00 182.5 F 0.0
20:00 179.4 F 0.0
25:00 179.4 F 0.0
30:00 167.5 F 0.0

Drying Complete

FILLING: 2:21
HEATING: 17:57
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:57:23
CYCLE COMPLETE

04/26/2021 12:44 PM
ACCEPTED: X
REJECTED:
Operator: U01