
Midmark M11 - v1.0.0
Total Cycles: 2454

Name: RITTER 1
Sterilizer ID: V2178198
10/19/2022 07:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.1 F 0.0
2:00 83.0 F 0.0
4:00 103.7 F 0.0
6:00 132.6 F 2.2
8:00 158.4 F 5.0
10:00 175.2 F 8.5
12:00 197.7 F 12.2
14:00 207.0 F 3.5
16:00 217.7 F 6.9
18:00 227.7 F 10.4
20:00 233.8 F 8.7
22:00 238.3 F 10.8
24:00 234.5 F 7.7
26:00 247.4 F 14.2
28:00 258.2 F 20.1
30:00 268.3 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.1 F 29.4
2:00 272.5 F 29.1
3:00 271.3 F 28.5
4:00 271.5 F 29.0

Min 271.1 F 28.4
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
5:00 216.3 F 0.0
10:00 212.7 F 0.0
15:00 214.4 F 0.0
20:00 211.8 F 0.0
25:00 217.7 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:23
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:09:27
CYCLE COMPLETE

10/19/2022 08:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____