

-----  
Midmark M11 - v1.0.0  
Total Cycles: 913

Name: RITTER 1  
Sterilizer ID: V2178198  
06/17/2021 04:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 132.1 F 0.0  
2:00 134.9 F 1.3  
4:00 150.0 F 3.8  
6:00 178.4 F 7.6  
8:00 199.3 F 12.2  
10:00 216.6 F 6.0  
12:00 225.6 F 11.1  
14:00 229.2 F 7.5  
16:00 242.4 F 13.5  
18:00 241.4 F 11.7  
20:00 254.3 F 18.1  
22:00 265.5 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.8 F 29.6  
2:00 271.7 F 28.8  
3:00 272.7 F 29.8  
4:00 272.4 F 29.3

Min 271.4 F 28.8  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.4 F 0.0  
5:00 223.2 F 0.0  
10:00 230.5 F 0.0  
15:00 227.1 F 0.0  
20:00 226.2 F 0.0  
25:00 227.3 F 0.0  
30:00 229.8 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 23:46  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 01:01:45  
CYCLE COMPLETE

06/17/2021 05:05 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_