

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4031

Name: RITTER2  
Sterilizer ID: V2313228  
10/16/2024 10:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.7 F 0.0  
2:00 146.2 F 1.4  
4:00 155.8 F 4.3  
6:00 181.1 F 8.3  
8:00 201.0 F 12.5  
10:00 214.5 F 4.8  
12:00 225.4 F 8.8  
14:00 235.7 F 13.2  
16:00 237.1 F 10.4  
18:00 236.5 F 9.0  
20:00 248.7 F 15.1  
22:00 259.7 F 21.3  
24:00 269.8 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.9 F 29.8  
2:00 272.0 F 29.1  
3:00 271.6 F 29.3  
4:00 272.8 F 29.8

Min 271.4 F 28.9  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.2 F 0.0  
5:00 197.0 F 0.0  
10:00 195.7 F 0.0  
15:00 193.3 F 0.0  
20:00 180.3 F 0.0  
25:00 179.4 F 0.0  
30:00 174.0 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 25:01  
STERILIZING: 4:00  
VENTING: 3:51  
DRYING: 30:00  
TOTAL CYCLE: 01:03:51  
CYCLE COMPLETE

10/16/2024 11:31 AM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_