
Midmark M11 - v1.0.0
Total Cycles: 3732

Name: RITTER 1
Sterilizer ID: V2178198
09/07/2023 10:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.6 F 0.0
2:00 186.9 F 1.4
4:00 196.2 F 5.1
6:00 207.8 F 9.9
8:00 227.3 F 7.2
10:00 232.9 F 9.5
12:00 239.8 F 10.2
14:00 242.1 F 12.1
16:00 240.4 F 10.8
18:00 253.2 F 17.0
20:00 263.9 F 23.3
22:00 272.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.6 F 29.0
2:00 271.2 F 28.3
3:00 272.6 F 29.1
4:00 271.9 F 28.5

Min 271.1 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.5 F 0.0
5:00 222.4 F 0.0
10:00 235.2 F 0.0
15:00 228.0 F 0.0
20:00 226.8 F 0.0
25:00 228.3 F 0.0
30:00 226.9 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 22:10
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:01:15
CYCLE COMPLETE

09/07/2023 11:25 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____