

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1985

Name: RITTER 1  
Sterilizer ID: V2178198  
06/01/2022 08:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 140.1 F 0.0  
2:00 142.1 F 2.3  
4:00 164.1 F 5.6  
6:00 185.0 F 9.9  
8:00 209.6 F 4.9  
10:00 218.6 F 7.2  
12:00 231.9 F 12.1  
14:00 232.6 F 8.4  
16:00 241.5 F 11.2  
18:00 244.6 F 12.6  
20:00 257.1 F 19.2  
22:00 268.3 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.4  
1:00 272.1 F 29.0  
2:00 272.1 F 28.7  
3:00 271.7 F 28.6  
4:00 272.5 F 29.1

Min 271.4 F 28.3  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.2 F 0.0  
5:00 211.7 F 0.0  
10:00 221.9 F 0.0  
15:00 207.1 F 0.0  
20:00 227.5 F 0.0  
25:00 235.7 F 0.0  
30:00 242.7 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 23:06  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:01:03  
CYCLE COMPLETE

06/01/2022 09:59 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_