
Midmark M11 - v1.0.5
Total Cycles: 2983

Name: RITTER2
Sterilizer ID: V2313228
10/05/2023 04:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.6 F 0.0
2:00 164.9 F 2.7
4:00 186.6 F 6.6
6:00 203.8 F 11.2
8:00 218.1 F 5.3
10:00 229.9 F 9.9
12:00 234.8 F 9.4
14:00 242.9 F 12.9
16:00 242.7 F 12.1
18:00 255.1 F 18.6
20:00 266.1 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.3 F 29.4
2:00 271.9 F 29.6
3:00 272.1 F 29.2
4:00 271.8 F 29.6

Min 271.3 F 28.9
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.2 F 0.0
5:00 205.5 F 0.0
10:00 209.5 F 0.0
15:00 207.6 F 0.0
20:00 204.7 F 0.0
25:00 201.7 F 0.0
30:00 199.0 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 21:38
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:00:55
CYCLE COMPLETE

10/05/2023 05:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____