

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2601

Name: RITTER2  
Sterilizer ID: V2313228  
07/04/2023 11:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.6 F 0.0  
2:00 182.5 F 3.6  
4:00 197.1 F 8.6  
6:00 216.6 F 9.4  
8:00 228.7 F 9.9  
10:00 234.6 F 8.2  
12:00 243.7 F 13.2  
14:00 243.9 F 12.7  
16:00 256.7 F 19.5  
18:00 268.0 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 272.5 F 29.7  
2:00 271.4 F 28.9  
3:00 272.4 F 29.9  
4:00 272.2 F 29.4  
  
Min 271.1 F 28.9  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.8 F 0.0  
5:00 217.8 F 0.0  
10:00 240.0 F 0.0  
15:00 232.8 F 0.0  
20:00 221.2 F 0.0  
25:00 219.5 F 0.0  
30:00 226.3 F 0.0  
Drying Complete

FILLING: 1:59  
HEATING: 19:17  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:28  
CYCLE COMPLETE

07/04/2023 12:05 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_