
Midmark M11 - v1.0.5
Total Cycles: 2115

Name: RITTER2
Sterilizer ID: V2313228
03/03/2023 09:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.5 F 0.0
2:00 80.8 F 0.0
4:00 88.4 F 0.0
6:00 120.8 F 0.0
8:00 157.2 F 3.8
10:00 187.1 F 8.0
12:00 206.0 F 12.8
14:00 219.3 F 6.6
16:00 231.3 F 11.0
18:00 233.0 F 7.2
20:00 244.5 F 13.4
22:00 243.9 F 12.2
24:00 256.7 F 19.1
26:00 268.2 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 271.9 F 28.9
2:00 272.3 F 29.2
3:00 272.5 F 29.4
4:00 272.7 F 29.4

Min 271.7 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 232.4 F 0.0
Dry Cycle Aborted

FILLING: 1:06
HEATING: 27:06
STERILIZING: 4:00
VENTING: 2:45
DRYING: 1:09
TOTAL CYCLE: 00:36:05

03/03/2023 10:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____