
Midmark M11 - v1.0.5
Total Cycles: 1540

Name: RITTER2
Sterilizer ID: V2313228
09/27/2022 08:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.3 F 0.0
2:00 166.0 F 1.9
4:00 170.2 F 5.6
6:00 193.0 F 10.6
8:00 212.9 F 4.8
10:00 226.9 F 10.3
12:00 230.7 F 7.9
14:00 241.9 F 13.0
16:00 241.8 F 11.9
18:00 254.4 F 18.4
20:00 265.7 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.4 F 29.9
2:00 271.4 F 29.4
3:00 272.3 F 30.0
4:00 271.7 F 29.3

Min 271.3 F 29.2
Max 272.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.1 F 0.0
5:00 206.5 F 0.0
10:00 219.3 F 0.0
15:00 224.4 F 0.0
20:00 226.9 F 0.0
25:00 229.7 F 0.0
30:00 204.8 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 21:45
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:46
CYCLE COMPLETE

09/27/2022 09:04 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____