
Midmark M11 - v1.0.5
Total Cycles: 1013

Name: RITTER2
Sterilizer ID: V2313228
04/20/2022 12:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.0 F 0.0
2:00 198.4 F 2.0
4:00 214.6 F 7.9
6:00 227.1 F 11.5
8:00 235.5 F 10.7
10:00 236.6 F 9.0
12:00 239.5 F 9.7
14:00 250.6 F 15.8
16:00 263.4 F 23.2
18:00 271.8 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.5 F 29.2
2:00 271.7 F 29.0
3:00 272.5 F 29.2
4:00 271.6 F 28.6

Min 271.6 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.5 F 0.0
5:00 212.2 F 0.0
10:00 224.4 F 0.0
15:00 229.1 F 0.0
20:00 230.6 F 0.0
25:00 228.1 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 2:33
HEATING: 18:03
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:48
CYCLE COMPLETE

04/20/2022 12:57 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____