

-----  
Midmark M11 - v1.0.5  
Total Cycles: 375

Name: RITTER2  
Sterilizer ID: V2313228  
09/16/2021 02:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.6 F 0.0  
2:00 157.9 F 2.9  
4:00 173.5 F 7.2  
6:00 195.7 F 12.6  
8:00 216.0 F 6.7  
10:00 231.2 F 12.1  
12:00 234.2 F 9.5  
14:00 238.2 F 9.4  
16:00 248.4 F 14.8  
18:00 260.4 F 21.7  
20:00 271.6 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.1  
1:00 272.5 F 29.3  
2:00 272.6 F 29.9  
3:00 272.2 F 29.0  
4:00 272.9 F 29.8

Min 271.7 F 28.9  
Max 273.1 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.3 F 0.0  
5:00 204.3 F 0.0  
10:00 209.5 F 0.0  
15:00 202.0 F 0.0  
20:00 202.1 F 0.0  
25:00 199.8 F 0.0  
30:00 200.6 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 20:34  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:59:05  
CYCLE COMPLETE

09/16/2021 03:33 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02