

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5304

Name: RITTER 1  
Sterilizer ID: V2178198  
01/31/2025 11:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.5 F 0.0  
2:00 169.0 F 2.3  
4:00 175.9 F 6.0  
6:00 195.5 F 10.3  
8:00 216.5 F 5.2  
10:00 222.6 F 6.7  
12:00 234.0 F 11.0  
14:00 232.6 F 7.0  
16:00 243.3 F 12.5  
18:00 240.8 F 10.7  
20:00 252.8 F 16.7  
22:00 263.4 F 22.9  
24:00 272.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.3  
1:00 272.1 F 29.1  
2:00 272.4 F 29.0  
3:00 271.8 F 28.6  
4:00 271.8 F 28.7

Min 271.0 F 28.2  
Max 273.0 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.9 F 0.0  
5:00 207.0 F 0.0  
10:00 217.3 F 0.0  
15:00 219.9 F 0.0  
20:00 218.0 F 0.0  
25:00 218.3 F 0.0  
30:00 218.6 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 24:22  
STERILIZING: 4:00  
VENTING: 3:56  
DRYING: 30:00  
TOTAL CYCLE: 01:03:19  
CYCLE COMPLETE

01/31/2025 12:05 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_