
Midmark M11 - v1.0.0
Total Cycles: 2369

Name: RITTER 1
Sterilizer ID: V2178198
09/26/2022 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.6 F 0.0
2:00 142.0 F 2.3
4:00 161.9 F 5.7
6:00 182.5 F 9.8
8:00 205.3 F 7.4
10:00 214.8 F 7.6
12:00 229.1 F 12.4
14:00 232.5 F 8.7
16:00 240.3 F 10.9
18:00 244.2 F 12.8
20:00 256.7 F 19.3
22:00 268.0 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.2 F 29.6
2:00 272.1 F 29.1
3:00 271.2 F 28.6
4:00 271.8 F 29.4

Min 271.1 F 28.6
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 210.4 F 0.0
10:00 230.8 F 0.0
15:00 221.1 F 0.0
20:00 229.5 F 0.0
25:00 227.4 F 0.0
30:00 229.3 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 23:22
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:01:33
CYCLE COMPLETE

09/26/2022 01:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____