
Midmark M11 - v1.0.5
Total Cycles: 386

Name: RITTER2
Sterilizer ID: V2313228
09/23/2021 06:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.0 F 0.0
2:00 73.4 F 0.0
4:00 79.0 F 0.1
6:00 101.0 F 0.1
8:00 145.3 F 1.6
10:00 170.5 F 4.4
12:00 189.6 F 7.7
14:00 200.8 F 11.1
16:00 217.6 F 5.7
18:00 225.5 F 9.8
20:00 235.1 F 10.2
22:00 237.6 F 11.8
24:00 238.5 F 10.3
26:00 251.1 F 16.5
28:00 261.1 F 22.6
30:00 270.9 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 30.0
1:00 271.6 F 30.1
2:00 271.5 F 30.0
3:00 271.5 F 30.0
4:00 271.9 F 30.0

Min 271.4 F 29.9
Max 273.4 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.6 F 0.0
5:00 237.2 F 0.0
10:00 227.4 F 0.0
15:00 224.6 F 0.0
20:00 226.1 F 0.0
25:00 221.9 F 0.0
30:00 216.4 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 31:12
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:09:36
CYCLE COMPLETE

09/23/2021 07:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____