
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 003705
10:12 27/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:36
STERILIZING 3:24
136.0 C 220kPa 3:24
136.4 C 220kPa 3:54
136.4 C 222kPa 4:24
136.4 C 224kPa 4:54
136.4 C 223kPa 5:24
136.4 C 220kPa 5:54
136.3 C 220kPa 6:24
136.4 C 220kPa 6:54
136.4 C 220kPa 6:54
VENTING 6:54
AIR DRYING 7:43
CYCLE COMPLETE 7:46

Digital Signature #
93AFAE7A67B71AE2
