

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1330

Name: RITTER 1  
Sterilizer ID: V2178198  
11/02/2021 09:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.2 F 0.0  
2:00 190.7 F 3.4  
4:00 198.8 F 8.4  
6:00 227.8 F 9.4  
8:00 229.5 F 10.0  
10:00 232.9 F 7.9  
12:00 244.1 F 12.5  
14:00 244.2 F 12.4  
16:00 256.9 F 19.2  
18:00 268.3 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 272.4 F 29.4  
2:00 271.7 F 28.7  
3:00 271.8 F 29.1  
4:00 272.2 F 29.1  
  
Min 271.0 F 28.5  
Max 273.0 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.9 F 0.0  
5:00 241.3 F 0.0  
10:00 237.5 F 0.0  
15:00 223.4 F 0.0  
20:00 189.5 F 0.0  
25:00 169.3 F 0.0  
30:00 179.0 F 0.0  
  
Drying Complete

FILLING: 1:43  
HEATING: 19:08  
STERILIZING: 4:00  
VENTING: 3:04  
DRYING: 30:00  
TOTAL CYCLE: 00:57:55  
CYCLE COMPLETE

11/02/2021 10:34 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01