

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5000

Name: RITTER 1  
Sterilizer ID: V2178198  
10/16/2024 01:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 154.3 F 0.0  
2:00 157.3 F 0.1  
4:00 171.0 F 0.7  
6:00 204.6 F 3.9  
8:00 207.5 F 8.0  
10:00 222.8 F 12.1  
12:00 229.1 F 7.2  
14:00 239.0 F 12.3  
16:00 237.1 F 9.7  
18:00 239.1 F 9.1  
20:00 248.3 F 14.4  
22:00 259.8 F 20.7  
24:00 270.3 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.4 F 29.1  
2:00 272.1 F 28.6  
3:00 271.1 F 28.2  
4:00 271.9 F 28.9

Min 271.1 F 28.1  
Max 272.5 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.7 F 0.0  
5:00 192.4 F 0.0  
10:00 189.7 F 0.0  
15:00 190.8 F 0.0  
20:00 215.0 F 0.0  
25:00 219.3 F 0.0  
30:00 224.9 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 24:56  
STERILIZING: 4:00  
VENTING: 3:50  
DRYING: 30:00  
TOTAL CYCLE: 01:03:37  
CYCLE COMPLETE

10/16/2024 02:12 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_