

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1032

Name: RITTER2  
Sterilizer ID: V2313228  
04/26/2022 08:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.3 F 0.0  
2:00 161.5 F 1.9  
4:00 167.6 F 5.1  
6:00 188.0 F 9.6  
8:00 210.3 F 6.5  
10:00 225.1 F 8.3  
12:00 238.4 F 13.4  
14:00 238.3 F 10.7  
16:00 236.0 F 9.1  
18:00 250.2 F 15.9  
20:00 262.1 F 22.6  
22:00 272.6 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.1  
1:00 272.1 F 29.2  
2:00 272.4 F 29.7  
3:00 271.4 F 29.1  
4:00 272.6 F 29.4

Min 271.4 F 28.9  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.5 F 0.0  
5:00 238.7 F 0.0  
10:00 223.1 F 0.0  
15:00 205.3 F 0.0  
20:00 187.5 F 0.0  
25:00 185.4 F 0.0  
30:00 184.8 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 22:22  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:00:41  
CYCLE COMPLETE

04/26/2022 09:36 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02