
Midmark M11 - v1.0.0
Total Cycles: 1874

Name: RITTER 1
Sterilizer ID: V2178198
04/26/2022 03:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.7 F 0.0
2:00 181.4 F 3.4
4:00 193.8 F 7.7
6:00 213.4 F 12.7
8:00 225.4 F 7.1
10:00 237.3 F 12.5
12:00 236.6 F 9.6
14:00 237.2 F 8.7
16:00 248.9 F 14.6
18:00 261.2 F 21.4
20:00 272.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.6 F 28.9
2:00 271.3 F 28.1
3:00 272.5 F 29.0
4:00 271.8 F 28.2

Min 271.2 F 28.0
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.4 F 0.0
5:00 220.1 F 0.0
10:00 205.5 F 0.0
15:00 215.0 F 0.0
20:00 210.8 F 0.0
25:00 212.3 F 0.0
30:00 218.9 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:01
CYCLE COMPLETE

04/26/2022 04:30 PM
ACCEPTED: X
REJECTED:
Operator: U02