
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000288
11:16 09/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:31
STERILIZING 3:34
136.0 C 227kPa 3:34
136.4 C 231kPa 4:04
136.4 C 233kPa 4:34
136.3 C 232kPa 5:04
136.4 C 233kPa 5:34
136.4 C 233kPa 6:04
136.3 C 232kPa 6:34
136.3 C 231kPa 7:04
136.3 C 231kPa 7:04
VENTING 7:04
AIR DRYING 7:57
CYCLE COMPLETE 15:07

Digital Signature #
63D777B03CEF2198
