
Midmark M11 - v1.0.0
Total Cycles: 3398

Name: RITTER 1
Sterilizer ID: V2178198
06/15/2023 02:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.3 F 0.0
2:00 180.0 F 0.0
4:00 198.9 F 4.3
6:00 218.0 F 9.0
8:00 228.0 F 11.3
10:00 233.0 F 9.9
12:00 237.7 F 8.9
14:00 244.9 F 13.3
16:00 243.6 F 12.2
18:00 256.1 F 18.7
20:00 267.1 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.5 F 29.0
2:00 271.5 F 28.2
3:00 272.5 F 29.1
4:00 272.1 F 28.6

Min 271.5 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 215.3 F 0.0
10:00 221.6 F 0.0
15:00 222.7 F 0.0
20:00 222.0 F 0.0
25:00 221.0 F 0.0
30:00 218.9 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 21:30
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 00:59:49
CYCLE COMPLETE

06/15/2023 03:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____