
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001513
12:04 18/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:08
PRESSURIZING 2:12
STERILIZING 3:08
136.4 C 223kPa 3:08
136.4 C 226kPa 3:38
136.2 C 223kPa 4:08
136.4 C 223kPa 4:38
136.4 C 227kPa 5:08
136.1 C 223kPa 5:38
136.4 C 226kPa 6:08
136.2 C 225kPa 6:38
136.2 C 225kPa 6:38
VENTING 6:38
AIR DRYING 7:31
CYCLE COMPLETE 7:40

Digital Signature #
EF0D072569BC3529
