

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1946

Name: RITTER2  
Sterilizer ID: V2313228  
01/18/2023 01:21 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.0 F 0.0  
2:00 164.7 F 3.3  
4:00 182.1 F 7.4  
6:00 205.5 F 12.5  
8:00 217.7 F 6.6  
10:00 233.6 F 11.8  
12:00 234.0 F 9.2  
14:00 238.7 F 9.7  
16:00 247.2 F 14.4  
18:00 258.9 F 21.1  
20:00 270.2 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.7  
1:00 272.9 F 30.3  
2:00 271.9 F 29.6  
3:00 273.0 F 30.5  
4:00 272.6 F 29.9

Min 271.8 F 29.6  
Max 273.0 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.7 F 0.0  
5:00 227.5 F 0.0  
10:00 234.5 F 0.0  
15:00 228.7 F 0.0  
20:00 230.4 F 0.0  
25:00 230.7 F 0.0  
30:00 228.1 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 21:44  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:00:07  
CYCLE COMPLETE

01/18/2023 02:21 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_