
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000951
14:01 09/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:59
PRESSURIZING 2:01
STERILIZING 3:04
136.4 C 220kPa 3:04
136.2 C 220kPa 3:34
136.4 C 225kPa 4:04
136.4 C 224kPa 4:34
136.4 C 222kPa 5:04
136.1 C 223kPa 5:34
136.2 C 224kPa 6:04
136.3 C 224kPa 6:34
136.3 C 224kPa 6:34
VENTING 6:34
AIR DRYING 7:26
CYCLE COMPLETE 7:43

Digital Signature #
3DD5BADAB648612B
