
Midmark M11 - v1.0.0
Total Cycles: 4495

Name: RITTER 1
Sterilizer ID: V2178198
04/26/2024 09:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.8 F 0.0
2:00 171.9 F 2.3
4:00 175.9 F 5.9
6:00 222.4 F 10.1
8:00 221.5 F 5.3
10:00 226.6 F 8.1
12:00 238.2 F 12.7
14:00 237.5 F 9.4
16:00 239.2 F 9.6
18:00 247.5 F 13.6
20:00 258.8 F 19.8
22:00 269.3 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.3
1:00 272.6 F 29.0
2:00 272.2 F 28.5
3:00 272.1 F 28.9
4:00 272.6 F 28.7

Min 271.4 F 28.2
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.9 F 0.0
5:00 222.4 F 0.0
10:00 223.8 F 0.0
15:00 230.0 F 0.0
20:00 226.2 F 0.0
25:00 230.6 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 23:47
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:02:54
CYCLE COMPLETE

04/26/2024 10:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____