

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4823

Name: RITTER 1  
Sterilizer ID: V2178198  
08/15/2024 03:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 92.8 F 0.0  
2:00 95.6 F 0.1  
4:00 105.6 F 0.5  
6:00 133.7 F 1.1  
8:00 176.3 F 1.3  
10:00 191.4 F 4.4  
12:00 210.0 F 7.8  
14:00 221.8 F 11.4  
16:00 224.6 F 5.1  
18:00 232.9 F 10.3  
20:00 236.0 F 8.6  
22:00 244.1 F 12.8  
24:00 242.3 F 11.3  
26:00 254.5 F 17.6  
28:00 264.8 F 23.7  
30:00 271.4 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.1  
1:00 271.8 F 28.7  
2:00 272.3 F 28.9  
3:00 272.0 F 28.5  
4:00 272.5 F 28.8

Min 271.0 F 28.0  
Max 273.0 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.4 F 0.0  
5:00 209.3 F 0.0  
10:00 199.4 F 0.0  
15:00 213.3 F 0.0  
20:00 228.9 F 0.0  
25:00 239.4 F 0.0  
30:00 236.5 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 30:07  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:08:40  
CYCLE COMPLETE

08/15/2024 04:20 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_