
Midmark M11 - v1.0.5
Total Cycles: 2913

Name: RITTER2
Sterilizer ID: V2313228
09/21/2023 08:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.9 F 0.0
2:00 71.0 F 0.0
4:00 71.9 F 0.0
6:00 89.2 F 0.0
8:00 146.2 F 0.1
10:00 177.5 F 0.7
12:00 199.1 F 3.5
14:00 210.2 F 6.7
16:00 219.6 F 10.1
18:00 230.8 F 10.1
20:00 232.1 F 9.4
22:00 240.0 F 10.9
24:00 241.8 F 11.9
26:00 239.7 F 10.4
28:00 252.2 F 16.7
30:00 262.9 F 22.9
32:00 272.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.6 F 29.0
2:00 272.0 F 29.2
3:00 271.9 F 28.7
4:00 272.7 F 29.4

Min 271.4 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.1 F 0.0
Dry Cycle Aborted

FILLING: 0:49
HEATING: 32:20
STERILIZING: 4:00
VENTING: 3:22
DRYING: 2:43
TOTAL CYCLE: 00:43:14

09/21/2023 09:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____