
Midmark M11 - v1.0.0
Total Cycles: 2002

Name: RITTER 1
Sterilizer ID: V2178198
06/07/2022 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.7 F 0.0
2:00 175.2 F 3.8
4:00 184.6 F 8.2
6:00 204.8 F 11.0
8:00 227.5 F 8.7
10:00 237.2 F 10.3
12:00 240.2 F 11.2
14:00 237.3 F 9.2
16:00 251.5 F 16.0
18:00 263.2 F 22.8
20:00 271.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.6
1:00 272.6 F 29.3
2:00 271.4 F 28.5
3:00 272.6 F 29.3
4:00 271.5 F 28.6

Min 271.2 F 28.5
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.5 F 0.0
5:00 229.5 F 0.0
10:00 227.6 F 0.0
15:00 228.0 F 0.0
20:00 222.2 F 0.0
25:00 225.7 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 20:43
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:15
CYCLE COMPLETE

06/07/2022 01:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____