

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2870

Name: RITTER2  
Sterilizer ID: V2313228  
09/12/2023 12:33 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.9 F 0.0  
2:00 161.0 F 3.0  
4:00 182.5 F 7.0  
6:00 204.1 F 12.0  
8:00 223.0 F 7.1  
10:00 236.5 F 12.5  
12:00 236.7 F 9.9  
14:00 237.0 F 9.0  
16:00 249.3 F 15.1  
18:00 260.9 F 21.9  
20:00 271.6 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.4  
1:00 272.9 F 29.8  
2:00 271.4 F 29.5  
3:00 273.3 F 29.9  
4:00 271.4 F 29.2  
  
Min 271.3 F 29.2  
Max 273.3 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.0 F 0.0  
5:00 241.1 F 0.0  
10:00 196.2 F 0.0  
15:00 194.1 F 0.0  
20:00 191.3 F 0.0  
25:00 188.9 F 0.0  
30:00 185.7 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 21:22  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:00:33  
CYCLE COMPLETE

09/12/2023 01:33 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_