
Midmark M11 - v1.0.0
Total Cycles: 4811

Name: RITTER 1
Sterilizer ID: V2178198
08/13/2024 02:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.2 F 0.0
2:00 121.3 F 0.2
4:00 145.3 F 1.2
6:00 173.6 F 3.6
8:00 207.3 F 7.2
10:00 226.6 F 10.8
12:00 225.5 F 5.7
14:00 228.5 F 8.4
16:00 239.1 F 12.7
18:00 235.8 F 9.0
20:00 242.4 F 11.1
22:00 244.3 F 12.2
24:00 255.5 F 18.0
26:00 265.4 F 24.0
28:00 271.1 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.2
1:00 272.1 F 28.8
2:00 272.3 F 28.8
3:00 271.8 F 28.8
4:00 272.3 F 28.9

Min 270.9 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.6 F 0.0
5:00 200.0 F 0.0
10:00 192.5 F 0.0
15:00 211.2 F 0.0
20:00 170.8 F 0.0
25:00 170.8 F 0.0
30:00 172.2 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 28:05
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:06:36
CYCLE COMPLETE

08/13/2024 03:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____