
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000325
10:27 11/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:16
STERILIZING 4:33
136.2 C 230kPa 4:33
136.4 C 233kPa 5:03
136.3 C 233kPa 5:33
136.1 C 231kPa 6:03
136.3 C 230kPa 6:33
136.4 C 233kPa 7:03
136.1 C 232kPa 7:33
136.3 C 233kPa 8:03
136.3 C 233kPa 8:03
VENTING 8:04
AIR DRYING 8:59
CYCLE COMPLETE 16:01

Digital Signature #
76E4F1CD9C561558
