
Midmark M11 - v1.0.0
Total Cycles: 1807

Name: RITTER 1
Sterilizer ID: V2178198
04/11/2022 08:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.2 F 0.0
2:00 191.4 F 2.8
4:00 191.4 F 7.1
6:00 208.1 F 12.0
8:00 226.4 F 5.5
10:00 237.2 F 10.5
12:00 232.1 F 7.5
14:00 245.6 F 13.1
16:00 242.9 F 11.4
18:00 255.5 F 17.9
20:00 266.6 F 24.6
22:00 272.8 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.1 F 28.4
2:00 272.6 F 28.7
3:00 272.3 F 28.9
4:00 271.4 F 28.3

Min 271.4 F 28.3
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.1 F 0.0
5:00 225.0 F 0.0
10:00 214.3 F 0.0
15:00 212.9 F 0.0
20:00 224.9 F 0.0
25:00 231.9 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 22:15
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:50
CYCLE COMPLETE

04/11/2022 09:44 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____