
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005022
10:10 31/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:36
PRESSURIZING 3:36
STERILIZING 4:24
136.0 C 220kPa 4:24
136.3 C 220kPa 4:54
136.3 C 223kPa 5:24
136.2 C 224kPa 5:54
136.2 C 223kPa 6:24
136.3 C 222kPa 6:54
136.4 C 221kPa 7:24
136.3 C 220kPa 7:54
136.3 C 220kPa 7:54
VENTING 7:54
AIR DRYING 8:53
CYCLE COMPLETE 16:06

Digital Signature #
9B7DED980CD70DA1
