
Midmark M11 - v1.0.5
Total Cycles: 2033

Name: RITTER2
Sterilizer ID: V2313228
02/09/2023 02:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.9 F 0.0
2:00 187.2 F 3.0
4:00 190.6 F 7.4
6:00 218.0 F 13.1
8:00 229.1 F 8.9
10:00 236.6 F 9.9
12:00 243.8 F 13.2
14:00 243.7 F 12.7
16:00 257.5 F 19.9
18:00 269.5 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.2 F 29.5
2:00 271.4 F 29.1
3:00 272.3 F 29.6
4:00 271.6 F 28.9

Min 271.3 F 28.8
Max 272.4 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.8 F 0.0
5:00 212.3 F 0.0
10:00 228.7 F 0.0
15:00 229.7 F 0.0
20:00 221.8 F 0.0
25:00 224.5 F 0.0
30:00 228.4 F 0.0

Drying Complete

FILLING: 1:38
HEATING: 18:58
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:49
CYCLE COMPLETE

02/09/2023 03:22 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____