

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004310  
11:58 07/09/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:24  
PRESSURIZING 2:32  
STERILIZING 3:24  
136.0 C 220kPa 3:24  
136.2 C 223kPa 3:54  
136.2 C 221kPa 4:24  
136.1 C 223kPa 4:54  
136.2 C 222kPa 5:24  
136.4 C 224kPa 5:54  
136.4 C 224kPa 6:24  
136.4 C 224kPa 6:54  
136.4 C 224kPa 6:54  
VENTING 6:54  
AIR DRYING 8:49  
CYCLE COMPLETE 14:07

-----  
Digital Signature #  
D30B8C05269737FF  
-----