
Midmark M11 - v1.0.0
Total Cycles: 1038

Name: RITTER 1
Sterilizer ID: V2178198
07/21/2021 11:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.2 F 0.0
2:00 174.5 F 3.0
4:00 192.4 F 6.3
6:00 210.8 F 10.9
8:00 218.7 F 5.1
10:00 231.1 F 10.2
12:00 233.9 F 8.4
14:00 243.3 F 12.8
16:00 242.1 F 11.3
18:00 254.2 F 17.6
20:00 265.0 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 29.1
2:00 271.8 F 28.4
3:00 272.0 F 28.9
4:00 272.6 F 29.1

Min 271.5 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 209.8 F 0.0
10:00 203.7 F 0.0
15:00 199.6 F 0.0
20:00 198.6 F 0.0
25:00 199.8 F 0.0
30:00 194.5 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 21:58
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:00:53
CYCLE COMPLETE

07/21/2021 12:35 PM
ACCEPTED: X
REJECTED:
Operator: U01